



TECHNICAL SPECIFICATION AGAINST ENQUIRY NO 2761428E/ DT: 28.01.2017

Requirement of Multi Channel Cardial Analyser-----02 Numbers as per specification mentioned below:-

1. 12 CHANNEL ECG MACHINE with 12 lead simultaneous.
2. 5 to 7 inch Display Screen
3. Key board-Alphanumeric/patient information
4. Built-in ECG parameters measurements and interpretation
5. Internal memory minimum 100 ECG
6. Support External Archiving: Reteviewing of memory/Creating backup like USB/PC
7. Built-in rechargeable lithium ion battery
8. Recording –A4 size Z fold or Roll paper

PRE-QUALIFICATION CRITERIA FOR ECG MACHINE:

1. The vendor would have supplied at least two number of such type of equipment in the past one year.
2. The vendor has to submit the performance certificate from their customers to whom they have supplied such type of equipment.