



Enquiry Number 2770784E/10.08.2017

TECHNICAL SPECIFICATION FOR DIGITAL MANOMETER

- | | |
|------------------------|---|
| 1. Item Description | : Hand held Digital Micro Manometer |
| 2. Range &Resolution | : 0-140 mbar& 0.1mbar Resolution |
| 3. Display | : LCD |
| 4. Accuracy | : +/- 0.2% Full Scale |
| 5. Scale | : Manually selectable= mbar,mmWC |
| 6. Mode of Measurement | : Fast and Slow |
| 7. Power | : Battery Operated |
| 8. Test Certificate | : Calibration Certificate traceable to
National Standard |
| 9. O&M Manuals | : To be Furnished with Supply |
| 10. Guarantee | : 12 months from the date of delivery |

Total Number of Micro Manometer required: 10 (Ten)



Enquiry Number 2770784E/10.08.2017

QUALIFYING CRITERIA

1. Only those vendors who have supplied at least ten Digital Manometer for the past three years (from the date of opening of Tender) and such instrument is presently working satisfactorily for more than one year after delivery (from the date of opening of Tender) should quote.

The vendor should submit the following information where similar system has been supplied to qualify their offer.

2. Name and postal address of the customer or company where similar instrument is supplied.
 - a. Name and designation of the contact person of the customer
 - b. Phone, FAX no and email address of the contact person of the customer
 - c. Month and Year of supply of the instrument
 - d. Application for which the instrument is supplied
 - e. Along with the Technical offer, the Vendor should submit one Performance certificate from the customer for the satisfactory performance of the instrument supplied to them.