



An ISO 9001
Company

Bharat Heavy Electricals Limited

(High Pressure Boiler Plant)

Tiruchirappalli – 620014, TAMIL NADU, INDIA

MATERIALS MANAGEMENT / CAPITAL EQUIPMENT

ENQUIRY – Corrigendum 1

Phone: +91 431 257 7653
Email : skaruna@bheltry.co.in
spsey@bheltry.co.in
kmk@bheltry.co.in
Web : www.bhel.com

NIT No. 32752	Enquiry Number:	Enquiry Date:	Due date for submission of quotation:
	2601700002	12.05.2017	14.06.2017

You are requested to quote the Enquiry number date and due date in all your correspondences. This is only a request for quotation and not an order

Please note that under any circumstances both **delayed offer** and **late offers** will not be considered. Hence vendors are requested to ensure that the offer is reaching physically our office before 14.00 hrs on the date of tender opening

Item Description: X R D Stress Analyser

Details of Corrigendum

AMENDMENT TO CLAUSE Nos.3.5, 4.3 & 10.1 of PART - B


SI No of the Specification	Particulars / Description	BHEL Requirement (Existing)	BHEL Requirement (New) Amended as
3.5 of PART B	X R D STRESS ANALYZER	The goniometer shall have programmable provision for selection of ψ angles over a range of -60° to $+60^\circ$ minimum for measurement in multi exposure $\sin^2(\psi)$ and omega methods	The goniometer shall have programmable provision for selection of ψ angles over a range of -45° to $+45^\circ$ minimum for measurement in multi exposure $\sin^2(\psi)$ and omega methods.
4.3 of PART B	X-RAY TUBES	The tubes shall be suitable for measurement on the inner diameter side of the small diameter pipes	The X ray tubes shall be suitable for measurement on the inner diameter side of the large diameter pipes (>1000 mm).
10.1 of PART B	FIELD MEASUREMENT	Field measurement stand to enable x,y,z movements including tripod with arm arrangements	Suitable stand arrangement to enable X, Y, Z movement of the goniometer.

All other terms and conditions and specifications as published in the NIT 32752 remain unaltered.

BHEL's General guidelines / instructions (refer MM / CE / GENL / 001 - EMD) including bank guarantee formats and list of consortium banks, commercial terms compliance form can be downloaded from BHEL web site <http://www.bhel.com> or from the Government tender website <http://tenders.gov.in> (public sector units > Bharat Heavy Electricals Limited page) under above Enquiry reference.

Tenders should reach us before 14:00 hours on the due date
Tenders will be opened at 14:30 hours on the due date
Tenders would be opened in presence of the tenderers who have submitted their offers and who may like to be present

Yours faithfully,
For BHARAT HEAVY ELECTRICALS LIMITED


DGM / MM / Capital Equipment