



An ISO 9001
Company

Bharat Heavy Electricals Limited

(High Pressure Boiler Plant)

Tiruchirappalli – 620014, TAMIL NADU, INDIA

CAPITAL EQUIPMENT / MATERIALS MANAGEMENT

ENQUIRY

NOTICE INVITING TENDER

Phone: +91 431 257 7653

Email : skaruna@bheltry.co.in

Web : www.bhel.com

TWO PART BID

Tender to be submitted in two Parts

Enquiry
Number:

2631700054

Enquiry Date:

09.10.2017

Due date for submission of
quotation:

31.10.2017

You are requested to quote the Enquiry number date and due date in all your correspondence. This is only a request for quotation and not an order.

Please note that under any circumstances both delayed offer and late offers will not be considered. Hence vendors are requested to ensure that the offer is reaching physically our office before 14.00 hrs on the Date of tender opening.

Item	Description	Quantity
10	ECG Machine as per the technical specification & commercial conditions applicable (to be downloaded from web site www.bhel.com or https://eprocure.gov.in/epublish/app)	5 Nos.


Important points to be taken care during submission of offer

1. Delivery terms shall be FOR BHEL Stores, Trichy
2. Compliance Form No: MM/CE/TRY/ IND/ 08 to be filled and enclosed along with the offer failing which, the offer will not be considered for evaluation.
3. Delivery period required is 2 months from the date of Purchase Order.
4. Erection and commissioning shall be done at BHEL Trichy by the supplier
5. Time period required for Erection and Commissioning shall be 1 week from the date of intimation by BHEL
6. All updates, amendments, corrigenda, etc., (if any), for each tender will be posted only on the above websites from time to time, as and when required, until each tender is opened. There will be no publication of such updates, amendments, corrigenda, etc., through newspapers or any other media.

BHEL's General guidelines / instructions (refer MM/CE/GENL/001), compliance form and technical specification can be downloaded from BHEL web site <http://www.bhel.com> or from <https://eprocure.gov.in/epublish/app> under Enquiry reference "2631700054".

Tenders should reach us before 14:00 hours on the due date
Tenders will be opened at 14:30 hours on the due date
Tenders would be opened in presence of the tenderers who have submitted their offers and who may like to be present

Yours faithfully,
For BHARAT HEAVY ELECTRICALS LIMITED


DGM / Capital Equipment / MM

S. KARUNANIDHY

Dy. General Manager

Capital Equipment / MM

BHEL, Tiruchirappalli - 620 014.

ELECTROCARDIOGRAPHY (ECG) MACHINE, Qty. 5 nos.

Technical specification

S No	BHEL Specification	Bidder's offer
1	ECG machine type - 3 channel	
2	Simultaneous 3 channel ECG recording with 12 lead simultaneous acquisition for both adult and pediatric patients.	
3	The offered equipment should have a real time digital display of 3 channel ECG	
4	Modes of operation - Automatic, Manual & Rhythm.	
5	The offered equipment should have defibrillation protection	
6	Should have visual alarm for open lead	
7	Should have built in ECG parameters and interpretation	
8	Digital filtering to remove interference with automatic baseline correction.	
9	Should have a maintenance free digital thermal array printer	
10	Recording/writing speed should be 25mm/sec and 50mm/sec	
11	Printer should be able to print ECG on standard thermal paper roll. Bidder to specify the Length and width of the paper roll used. (The paper roll should be readily available in Local Market)	
12	Printer should have on / off selection	
13	Equipment should have sufficient battery (rechargeable type) backup for taking minimum 25 ECGs without AC power.	
14	The equipment should be compact and portable	
15	Weight of the equipment to be specified	
16	Should operate on mains AC 220V, 50Hz	
17	Vendor to provide the list of customers in Tamil Nadu or anywhere in India to whom such equipment has been supplied. Complete contact details of the customers to be provided.	
18	Service centre locations with complete address details to be provided	
19	The equipment shall be supplied with the latest software	
20	The vendor should have a valid CE / FDA certification	
21	Operation and Maintenance manual in English Language – 1 set.	

S No	BHEL Specification	Bidder's offer
22	Should be supplied with the following:- a) Patient cable – 2 nos. b) Chest electrodes (set of six) - 2 sets c) Limb electrodes (set of four) - 2 sets	
23	The installation of the equipment at BHEL hospital by supplier's representative.	
24	The equipment shall be guaranteed for a period of 12 months from the date of commissioning at BHEL works or 18 months from the date of despatch whichever is earlier.	