



An ISO 9001
Company

Bharat Heavy Electricals Limited
(High Pressure Boiler Plant)
Tiruchirappalli – 620014, TAMIL NADU, INDIA
CAPITAL EQUIPMENT / MATERIALS MANAGEMENT

ENQUIRY		Phone: +91 431 257 5443 / 7421	
NOTICE INVITING TENDER		Email : smams@bhel.in, vkaruna@bhel.in	
		Web : www.bhel.com	
TWO PART BID	Enquiry Number:	Enquiry Date:	Due date for submission of quotation:
Tender to be submitted in two Parts	2632100011	11.02.2021	10.03.2021
You are requested to quote the Enquiry number date and due date in all your correspondence. This is only a request for quotation and not an order. Please note that under any circumstances both delayed offer and late offers will not be considered. Hence vendors are requested to ensure that the offer is reaching physically our office before 14:00 hrs on the Date of tender opening.			
Item	Description		Quantity
10	Wax Bath as per the technical specification & commercial conditions applicable (to be downloaded from web site www.bhel.com or https://eprocure.gov.in/epublish/app)		1 No.
Important points to be taken care during submission of offer 1. Delivery terms shall be FOR, BHEL, Stores, High Pressure Boiler Plant, Tiruchirappalli - 620014. 2. Compliance Form TRY/ IND/ 04A to be filled and enclosed along with the offer failing which, the offer will not be considered for evaluation. 3. Delivery period required is Two Months from the date of Purchase Order. 4. Erection and commissioning activities shall be done by the supplier, at BHEL, Trichy works. 5. Time period required for Erection and Commissioning activities shall be 1 week from the date of intimation by BHEL 6. Procurement of the item through this NIT will be as per Public Procurement (Preference to Make in India), Order 2017 dated 04.06.2020 by DPIIT (DPIIT Order No. P-45021/2/2017-PP (BE-II) dated 04.06.2020), to encourage 'Make in India ' and promote manufacturing and production of goods and services in India. 7. All updates, amendments, corrigenda, etc., (if any), for each tender will be posted only on the above websites from time to time, as and when required, until each tender is opened. There will be no publication of such updates, amendments, corrigenda, etc., through newspapers or any other media. BHEL's General guidelines / instructions (refer MM/CE/GENL/001- NO EMD), compliance form and technical specification can be downloaded from BHEL web site http://www.bhel.com or from https://eprocure.gov.in/epublish/app under Enquiry reference "2632100011".			
Tenders should reach us before 14:00 hours on the due date Tenders will be opened at 14:30 hours on the due date Tenders would be opened in presence of the tenderers who have submitted their offers and who may like to be present		Yours faithfully, For BHARAT HEAVY ELECTRICALS LIMITED  S. MUTHUSUBRAMANIAN Manager Capital Equipment/MM BHEL, TRICHY - 620 014.	

WAX BATH, QTY. 1 NO.

Part A

S. No.	PARTICULARS	BIDDER's RESPONSE
1.1	Only those Bidders /OEMs should quote, who have supplied and commissioned at least 2 numbers of Wax bath in the past 5 years (as on the initial date of opening of Tender).	
1.2	List of customers to whom Wax bath were supplied, installed and commissioned till date to be provided	
1.3	Year of Launch of the Model quoted against this enquiry	
1.4	Details on SERVICE-AFTER-SALES Set-Up in India including the Addresses of Agents / Service Centres in South India.	
1.5	BHEL reserves the right to verify the information provided by the vendor. In case the information provided by the vendor is found to be false/ incorrect, the offer shall be rejected.	

Part B - Technical Specification

S NO	BHEL SPECIFICATION	BIDDER'S OFFER with technical details
1	Purpose: The equipment with tank, heating arrangement and trolley fitted with wheels is required for heating and maintaining molten wax for performing heat therapy in physiotherapy treatment.	
2	Should have stainless steel inner tank/mild steel powder coated.	
3	Tank capacity shall be suitable to hold minimum 30 kg wax	
4	Minimum wax that can be heated. Bidder to specify	
5	A removable insulated type lid/top cover for the tank with lifting handle shall be provided	
6	Suitable Industrial Grade Electrical heater to provided. Bidder to specify the details of the heating elements and rating in watts	
7	Thermostat manual control 30 – 90 Deg C or better with auto cut off to control wax temperature	
8	Power/Heater/thermostat ON indicator. Bidder to confirm	
9	Safety fuse. Bidder to specify	

S NO	BHEL SPECIFICATION	BIDDER'S OFFER with technical details
10	The equipment shall be properly insulated to avoid any electric shock to the operator. Bidder to confirm	
11	Weight in kg. Bidder to specify	
12	Overall Dimensions of the equipment (LxWxH) in mm. Bidder to specify	
13	Input power cable: length around 2 m. Bidder to specify	
14	The entire system to be mounted on trolley fitted with 4 wheels. Bidder to confirm.	
15	Bidder to specify the make and model of the offered equipment	
16	Operation & Maintenance manuals to be supplied in English. Hard copy – 1 set	
17	Input Power supply: 220-240 V AC (Single Phase) shall be provided by BHEL at one point.	
18	The test certificate / factory inspection test certificate of the equipment shall be submitted to BHEL for despatch clearance	
19	The equipment shall meet to the IEC-60601 standard Or applicable standard for safety, performance of medical electrical equivalent	
20	Installation, commissioning of the equipment shall be done by bidder's representative at BHEL hospital.	
21	The equipment shall be guaranteed for a period of 12 months from the date of commissioning or 18 months from the date of despatch whichever is earlier.	