

**TSGENCO
4 X 270 MW BHADRADRI TPS**

VOLUME II B

**TECHNICAL SPECIFICATION
FOR
ELECTRICAL LAB EQUIPMENT (CALIBRATION & DIAGNOSTICS)**

SPECIFICATION NO. : PE-TS-411-556-E001, REV. 03



**BHARAT HEAVY ELECTRICALS LIMITED
POWER SECTOR PROJECT ENGINEERING MANAGEMENT
NOIDA, 201301**

666070/2022/PS-PEM-EL



**TECHNICAL SPECIFICATION
FOR ELECTRICAL LAB EQUIPMENT
(CALIBRATION & DIAGNOSTIC)**

Doc. No. PE-TS-411-556-E001

Volume

Section

IIB


CONTENTS

Rev. : 03 DATE-01.02.2022

CONTENTS

<u>S. NO.</u>	<u>DESCRIPTION</u>	<u>NO. OF SHEETS</u>
01	SECTION – I	
a)	COMPLIANCE CERTIFICATE	01
b)	SPECIFIC TECHNICAL REQUIREMENTS	01
c)	ANNEXURE-I OF SECTION-I (DETAILED TECHNICAL PARAMETERS)	01
02	SECTION – II	
a)	STANDARD TECHNICAL SPECIFICATION	03
b)	ANNEXURE- I TO SECTION-II (DEMONSTRATION & HANDING OVER CHARGES)	01
c)	ANNEXURE-II OF SECTION-II (DATA SHEET-B FORMAT)	01
	 TOTAL NO. OF SHEETS= (INCLUDING COVER/ SEPARATOR SHEETS)	 08

666070/2022/PS-PEM-EL


	4X270 MW BHADRADRI TPS TECHNICAL SPECIFICATION FOR ELECTRICAL LABORATORY EQUIPMENTS (CALIBRATION & DIAGNOSTIC)	Doc. No. PE-TS-411-556-E001	
		Volume	IIB
		Section	I
	COMPLIANCE CERTIFICATE	Rev. : 03 DATE-01.02.2022	

COMPLIANCE CERTIFICATE

The bidder shall confirm compliance to the following by signing/ stamping this compliance certificate and furnishing same with the offer.

1. The scope of supply, technical details, construction features, design parameters etc. shall be as per technical specification & there are no exclusion/ deviation with regard to same.
2. There are no deviation with respect to specification other than those furnished in the 'schedule of deviations'
3. Only those technical submittals which are specifically asked for in NIT to be submitted at tender stage shall be considered as part of offer. Any other submission, even if made, shall not be considered as part of offer.
4. Any comments/ clarifications on technical/ inspection requirements furnished as part of bidder's covering letter shall not be considered by BHEL, and bidder's offer shall be construed to be in conformance with the specification.
5. Any changes made by the bidder in the price schedule with respect to the description/ quantities from those given in Annexure-A [BOQ-Cum-Price schedule] of the specification shall not be considered (i.e., technical description & quantities as per specification shall prevail).
6. All essential accessories for each electrical laboratory equipment required for testing various electrical equipment /devices during Commissioning, Operation and Maintenance of power plant to be ensured by bidder in their offer.

 BIDDER'S STAMP & SIGNATURE

	4X270 MW BHADRADRI TPS TECHNICAL SPECIFICATION FOR ELECTRICAL LABORATORY EQUIPMENTS (CALIBRATION & DIAGNOSTIC)		Doc. No. PE-TS-411-556-E001
			Volume IIB
			Section I
	SPECIFIC TECHNICAL REQUIREMENTS		Rev. : 03 DATE-01.02.2022

1.0 SCOPE

- 1.1 Design, manufacture, inspection and testing at manufacturer's work, proper packing and delivery to site of Electrical Laboratory Equipment as mentioned in different sections of this specification.
- 1.2 General technical requirements of the Electrical Laboratory Equipment are indicated in Section – II. Project specific technical requirements/changes are listed in section-I.
- 1.3 Electrical Laboratory Equipment shall be supplied along with all essential accessories required for the successful operation of the equipment.
- 1.4 Detailed technical parameters of Electrical laboratory equipment are listed in Annexure – I of Section-I.
- 1.5 The requirement of section-1 shall prevail and govern in case of conflict between the corresponding requirements of Section-I and section-II.

2.0 BILL OF QUANTITIES

Quantity requirement shall be as per 'BOQ cum price schedule' enclosed in NIT.


3.0 SPECIFIC TECHNICAL REQUIREMENT

S.No.	Reference Clause No. of Section-II (if any)	Specific Requirement/Change

4.0 DRAWING & DOCUMENTS TO BE SUBMITTED

- 4.1 Following documents/drawings for each Electrical Laboratory Equipment shall be submitted after placement of order for BHEL & customer's approval in addition to the documents/drawings listed in clause No. 8.2 of section-II:

Sl. No.	Drawings/Document description	Drawings/Document Number
1	Technical Data Sheet	PE-V0-411-556-E001
2	Catalogue	PE-V0-411-556-E002

	4X270 MW BHADRADRI TPS TECHNICAL SPECIFICATION FOR ELECTRICAL LABORATORY EQUIPMENTS (CALIBRATION & DIAGNOSTIC)	Doc. No. PE-TS-411-556-E001	
		Volume	IIB
		Section	I
	Annexure-I of SECTION-I DETAILED TECHNICAL PARAMETERS	Rev. : 03 DATE-01.02.2022	

1. Primary Injection Kit: 1 No

Make: MEGGER, OMICRON, DOBLE, ISA OR EQUIVALENT

Purpose:	For Commissioning application requiring high current injection, fully automatic with display and trolley
Input Voltage:	0-240 V, 1 Ph.
Output current in Amps:	0- 4000A at 24 KVA
No of Ranges:	3 or above
Period in Seconds:	0- 999.99
Resolution in Seconds:	0.001
Accuracy:	0.1% + 2 Digits
CTs:	Standard class CT's of accuracy 0.5
Accessory	10 M length cable

666070/2022/PS-PEM-EL



**4X270 MW BHADRADRI TPS
TECHNICAL SPECIFICATION FOR ELECTRICAL
LABORATORY EQUIPMENTS (CALIBRATION &
DIAGNOSTIC)**

Doc. No. PE-TS-411-556-E001

Volume

IIB

Section

I

Rev. : 03 DATE-01.02.2022

SECTION-II

STANDARD TECHNICAL SPECIFICATION



TECHNICAL SPECIFICATION FOR ELECTRICAL LABORATORY (CALIBRATION & DIAGNOSTICS) EQUIPMENT

SPECIFICATION NO. PE-TS-411-556-E002

VOLUME II

SECTION II

REVISION 00

DATE: 01.09.2017

SHEET 1 of 3

1.0 TECHNICAL REQUIREMENTS

- 1.1 Basic technical requirements for Electrical Lab Equipment shall be as indicated in this section, in addition to technical requirements specified in Section-I.
- 1.2 The intent of specification is not to specify herein all the details of design and manufacture. However, the equipment shall confirm in all respect to high standards of design engineering and workmanship and shall be capable of performing in continuous commercial operation up to vendors guarantee.
- 1.3 The Bidder may note that the equipment range, rating, quantities as detailed herein, are the minimum requirement only. All accessories for the equipment not covered here, if necessary for satisfactory and trouble free operation of the equipment, shall be quoted by the Bidder.
- 1.4 The instrument shall be suitable for satisfactory operation at an ambient temperature from 0°C to 55°C.
- 1.5 The Analog instruments shall be provided with knife-edge pointer and anti – parallax mirror.
- 1.6 The Bidder to quote only 'one' make/model against each equipment best to suit specification requirement.
- 1.7 The instrument shall be suitable for hand held operation, rugged in construction and suitable for field use.
- 1.8 All the equipment components shall be procured from reputed manufacturers and make of equipment shall be subject to the approval of BHEL/ BHEL's Customer.
- 1.9 **Bidder to note that "In case any offered make / model becomes obsolete or is stopped manufacturing by manufacturer, next higher model of the same make may be considered for ordering / supply at contract stage, without any price implication. In such cases, bidder is required to furnish valid confirmation letter from OEM as proof of change of model (citing reason: obsolete technology or stopping of manufacturing with date of effect) and that the offered model is "technically equivalent or better".**

2.0 CODES AND STANDARDS

Some of the standards, which shall generally be followed, are listed below. Other applicable relevant standards for any component part, even if not covered in listed standards shall be followed.

- i) IS – 6103 Method of test for specific resistance (resistivity) of electrical insulating fluid.
- ii) IS – 6700 Requirements of general purpose Cathode Ray Oscilloscope.
- iii) IS – 722 Specification for AC electricity meters.
- iv) IS – 8143 Specification for plugs & keys for resistance boxes.
- v) IS – 6104 Method of test for interfacial tension of oil against water.
- vi) IEC-51 Direct acting indicating analogue electrical measuring instruments and their accessories.



TECHNICAL SPECIFICATION FOR ELECTRICAL LABORATORY (CALIBRATION & DIAGNOSTICS) EQUIPMENT

SPECIFICATION NO. PE-TS-411-556-E002

VOLUME II

SECTION II

REVISION 00

DATE: 01.09.2017

SHEET 2 of 3

vii) Any other relevant National/ International standards as mentioned in Section-II with technical specification.

3.0 DESIGN CRITERIA:

- 3.1 Electrical Laboratory Equipment shall be used for calibration & diagnostics of various electrical equipment/ devices during Commissioning, Operation and Maintenance of power plant.
- 3.2 The Equipment will be kept in a clean but hot, humid and tropical atmosphere when not in use. Equipment will be placed in dust laden, hot, humid atmosphere during its use.
- 3.3 For continuous operation at specified rating, temperature rise of various equipment/ components shall be limited to the permissible value stipulated in the relevant standards and this specification.

4.0 TEST REQUIREMENTS:

- 4.1 All equipment to be tested as per latest edition of relevant standards. Further, testing shall include verification of physical, functional & technical parameters of equipment & accessories, in line with approved datasheet.
- 4.2 The tests shall be carried out by the vendor at in-house lab/ OEM lab/ Third party govt. accredited lab. Charges for all these tests shall be deemed to be included in the bid price.
- 4.3 Vendor to furnish test report to BHEL for review/ acceptance. BHEL/BHEL's customer may witness the testing of equipment.

5.0 PACKING:

All equipment shall be properly packed in Galvanized sheet steel trunk/ box with proper lock & key arrangement except for equipment which are trolley mounted or are already available in rugged steel/wooden box packing. Further, any damage (reading error/calibration error/broken parts/missing parts etc) found on receipt at site, leading to replacement of parts/whole item, shall be to bidder's account.

6.0 DEMONSTRATION TO BHEL / BHEL'S CUSTOMER

- 6.1 The vendor shall be responsible for demonstration of the supplied equipment at site, conforming the satisfactory operation.
- 6.2 The equipment for which demonstration is required at site shall be intimated by BHEL.
- 6.3 The charges for visit to site for demonstration at site shall be in-line with Annexure- I of Section-II.

7.0 PERFORMANCE GUARANTEE

The bidder shall guarantee that the equipment offered shall meet the requirement as stipulated in this specification and as confirmed by them in Technical Data Sheet. In case the performance of equipment is not as per performance guarantee, the bidder will have to replace the equipment at site free of cost.



**TECHNICAL SPECIFICATION FOR
ELECTRICAL LABORATORY
(CALIBRATION & DIAGNOSTICS)
EQUIPMENT**

SPECIFICATION NO. PE-TS-411-556-E002

VOLUME II

SECTION II

REVISION 00

DATE: 01.09.2017

SHEET 3 of 3

8.0 DRAWING DATA & MANUAL

8.1 To be submitted with the bid as technical offer:

- a) Compliance to Technical parameters of various equipment as specified in Annexure-I of Section-I in Data Sheet – B (as enclosed).
- b) Technical leaflets/ Catalogues/Product Manual of the Equipment

The Bidder may note that the drawing, data and manual listed herein are minimum requirement only. The Bidder shall ensure that the other necessary write-ups, curves and information required to fully describe the equipment are submitted with the bid.

8.2 Following documents/drawings shall be submitted after placement of order for BHEL & customer's approval:


- a) Technical Datasheet for each equipment
- b) Technical leaflets/ Catalogues/Product Manual of the Equipment
- c) Tests Reports and calibration certificate
- d) General arrangement drawing showing constructional features, accessories, connections, range and rating, mounting arrangement, space requirement etc.
- e) Detail instructions for application, assembly & testing of equipment.
- f) Wiring and schematic diagrams (if applicable).
- g) Instruction manual/ O&M manual of individual equipment

8.3 Instruction manual/O&M Manual of individual equipment

The manual shall clearly indicate in English the installation and connection method, check list of the tests to be carried out before commissioning of equipment. Maintenance and Calibration method shall also be provided in the manual.

8.4 Bidder to furnish all user instruction manuals, maintenance, handling, installation manuals & all test reports complete in all respect in bound volumes & soft copies to BHEL / BHEL's customer at the time of handing over the same to BHEL / BHEL's Customer.

8.5 Bidder to note that quoted item cost shall include cost of main item, cost of all accessories required for successful operation of equipment and testing cost of all equipment test as per relevant standard.

	4X270 MW BHADRADRI TPS TECHNICAL SPECIFICATION FOR ELECTRICAL LABORATORY EQUIPMENTS (CALIBRATION & DIAGNOSTIC)	Doc. No. PE-TS-411-556-E001	
		Volume	IIB
		Section	II
	ANNEXURE-I OF SECTION – II DEMONSTRATION & HANDING OVER CHARGES	Rev. : 03 DATE-01.02.2022	

**SCHEDULE OF PRICES FOR DEMONSTRATION & HANDING OVER TO
BHEL / BHEL'S CUSTOMER**

SL. NO.	DETAILS	ACTIVITY	UNIT CHARGES
1	LUMP SUM ALL INCLUSIVE CHARGES PER VISIT FOR EXPERIENCED / CAPABLE ENGINEER (EXCEPT DAILY CHARGES)	1 VISIT	20000/-
2	LUMP SUM ALL INCLUSIVE CHARGES FOR EXPERIENCED / CAPABLE ENGINEER PER DAY	1 DAY	5000/-

Note:

- TOTAL CHARGES = (Charges as per S.No.1) + [No. of Days(*) x Unit Charges as per Sl. No. 2]
- *: To be certified by BHEL site
- Bidder to note that provision of maximum 1 visit amounting to total 4 man days is envisaged for respective bidders.

4X270 MW BHADRADRI TPS								
TECHNICAL SPECIFICATION FOR ELECTRICAL LABORATORY EQUIPMENTS								
PE-TS-411-556-E001, Rev.-03								
DATA SHEET-B (Annexure-II of Section-II)								
NOTE :-								
1) EQUIPMENT TO BE SUPPLIED ALONG WITH ESSNTIAL ACCESSORIES FOR SUCCESSFUL OPERATION OF EQUIPMENT AT SITE .i.e. CLAMPS, CLIPS,LEADS ,CARRYING CASE etc.								
2) ALL EQUIPMENTS SHALL BE SUPPLIED WITH VALID CALIBRATION CERTIFICATE, WHEREVER APPLICABLE.								
3) All equipment shall be properly packed in Galvanized sheet steel trunk/ box with proper lock & key arrangement except for equipment which are trolley mounted or are already available in rugged steel/wooden box packing. Further, any damage (reading error/calibration error/broken parts/missing parts etc) found on receipt at site, leading to replacement of parts/whole item, shall be to bidder's account								
Item description	Whether Quoted or not (Yes/No)	Make	Model no	Country of Manufacture	Authorizatio n letter (attached or not)	Bidder to indicate the parameters of offered model against each parameter mentioned in the tender specification so that it can be concluded that offered model is meeting all requiremnets as specified in technical specification	Catalogue (attached or not)	Remarks
1. Primary Injection Kit: 1 No								
Purpose: For Commissioning application requiring high current injection, fully automatic with display and trolley								
Input Voltage: 0-240 V, 1 Ph.								
Output current in Amps: 0- 4000A at 24 KVA								
No of Ranges: 3 or above								
Period in Seconds: 0- 999.99								
Resolution in Seconds: 0.001								
Accuracy: 0.1% + 2 Digits								
CTs: Standard class CT's of accuracy 0.5								
Accessory :10 M length cable								