

Annexure - B
Pre-Qualification Criteria

Item description: Standard Reference Impact specimen for calibration of Impact Testing machine as per ASTM E23

The bidder has to compulsorily meet all the following conditions and supporting documents are to be furnished, without which, the offer will not be considered for technical evaluation.
--

1. SRM (Standard Reference Material) conforming to ASTM E23 (latest revision) must be sourced from NIST, USA
--

2. Necessary documents / Certificates from NIST, USA (for each of the three SRM ID numbers) must be submitted along with the materials
--

3. The material will be accepted only after physical verification and its associated documents, after its receipt at BHEL, Trichy
