

CORRIGENDUM-2

Tender enquiry No. R/5599/2016/2125V1 dt: 09/07/2016 with tender opening date 30/07/2016 for the item "Ambient Air Quality Monitoring System (AAQMS) for Yeramarus TPP"

As a result of pre-bid meeting held on 26/07/2016, following clarification mentioned below should be read in conjunction with details and specification mentioned in original tender document. All bidders to quote accordingly.

Clause 2.4.2 of Specification : Equipment to be supplied with Inlet head for measurement of TSPM and PM₁₀ to be measured one at a time. TSPM head if desired (separately) be quoted as a part of price bid.

Analyzer for PM_{2.5} measurement to be provided as per clause 2.5 of product purchase specification.

Clause 2.5.5 of Specification : Average reading of 24 hrs is also acceptable.

Clause 2.7.6 Specification: Minimum detectable for multi gas calibrator shall be ignored.

Clause 2.9 of Specification: The required AAQMS shall be work as standalone unit and no Central Station is required. The quoted model should be equipped with wireless connectivity based on GSM/GPRS/RF modem or better to transfer data from AAQMS system to other location within the plant boundary of the Yeramarus thermal power station.

Necessary hardware, software for data transmission from AAQMS system will be in bidder's scope. Desired software at receiver end (other location within the plant boundary of the Yeramarus thermal power station) along-with necessary hardware for data reception will also be in bidder's scope.

Necessary computer at receiver end would not be in the scope of supply of the bidder.

Clause 2.9 of Specification: Bullet point 2 to be ignored.

Clause 2.10 Specification: Nitrogen cylinder with regulator is not required and to be ignored from the scope of supply.

Clause 2.13 of Specification: Product purchase specification will prevail.

5KVA capacity of UPS as mentioned in price bid format shall be ignored. Capacity of the UPS should be as per the requirement of quoted model.

Clause 2.14.1 of Specification: Thickness of outer sheet and inner sheet of the cabin shall be minimum 2.0 mm and 2.0 mm respectively. One (01) no. of door in the cabin to be provided. Flooring of cabin shall be provided with Aluminum checkered sheet of minimum 4.0 mm. The cabin walls and roof shall be insulation suitably of minimum 100 mm thickness.

Clause 2.14.3 (Point No 1) of Specification: Two (02) nos. of Air conditioners to be provided along with changeover unit.

Clause 2.14.3 (Point No 9 & 10) of Specification: Desired power supply at 01 no location at AAQMS would be provided other works as applicable in point 9 and 10 would be in bidder's scope.

Clause 2.15 & 2.16 of Specification: Product purchase specification will prevail.

Clause 4.7 of Specification: Consumable and essential spare for 18 months is to be quoted for future reference only and will not be considered during evaluation of present bid.

MSE provisions as per Govt. rules will prevail.

The Note-2 of the suggestive price bid format may please be read as "if any other component/system required for the operation of the quoted model the same should be indicated in the technical bid and the price for the same in price bid".

The last date to get the tender from BHEL has been extended from 29/07/2016 to 09/08/2016.

The part-1 bid opening date has been extended from 30/07/2016 to 10/08/2016.

FOR & ON BEHALF OF BHEL, HARIDWAR

Engineer (PCRI-BOI)