

CORRIGENDUM

Tender Enquiry No. R/5599/2019/3310V/1 Dt: 10/01/2020 for the procurement of item 'Laboratory TOC Analyser' for Maitree site

Corrigendum detail:

A. Following points may please be noted and covered by bidders:

- Analyser shall comply with all specifications according to standard Methods 5310B/5310C, EPA415.1/415.2/415.3, EPA SW-846 Test Method 9060/9060A in place of "Analyser shall comply with all specifications according to standard Methods 5310C, EPA415.2, EPA SW-46 Test Method 9060A".

B. Due date of submission & opening of subject tender has been extended up to 15/02/2020.

4 Singh
08/02/2020

4 Singh
08/02/2020