

Corrigendum-01 dated 22.05.2020 for

P.I. No. 241506350

(Portable type Silt / Suspended solid content measuring equipment)

PROJECT: BAIRASIUL HEP & MANGDECHHU HEP

REF. SPECIFICATION NO. : 241506350-1 & 241506350-2


The point no. 3 of clause no. 2.1 under heading "2. Features" of above specifications is as follows:

3. The silt / suspended solid contents in the water at site, that instrument will have to measure, shall be up to 10,000 mg/l. However, if the instrument can measure contents above 10,000 mg/l, the same shall be acceptable subject to maximum value of 60,000 mg/l.

However, the above clause of the specification may please be read as follows:

3. The silt / suspended solid contents in the water at site, that instrument will have to measure, shall be up to 10,000 mg/l. However, if the instrument can measure contents above 10,000 mg/l, the same shall also be acceptable.


दलपत बोराणा / DALPAT BORANA
उप प्रबंधक / Dy. Manager
एच.पी.ई. विभाग / HPE Division
बी.एच.ई.एल., भोपाल / BHEL, Bhopal


विशाल कुमार नामदेव / V.K. NAMDEO
प्रबंधक / Manager
एच.पी.ई. विभाग / HPE Division
बी.एच.ई.एल., भोपाल / BHEL, Bhopal


मनीष साहू / MANISH SAHU
उप महाप्रबंधक / Dy. General Manager
एच.पी.ई. विभाग / HPE Division
बी.एच.ई.एल., भोपाल / BHEL, Bhopal