

**TENDER ENQUIRY OF “FQA LAB – ELECTRICAL”
FOR 3x800 MW PVUNL PATRATU TPP PHASE I**

Ref No.: PE/PG/PA1/E- 6802/2021

Date:18.11.2021

CORRIGENDUM/ ADDENDUM- 1

Tender Enquiry no.: PE/PG/PA1/E- 6802/2021; Date: 08/11/2021,

AMMENDMENT 1 TO TECHNICAL SPECIFICATION

MODIFIED CLAUSES AS BELOW.

Sr. No.	Vol. No.	Section/ Description	Clause no	Page no	Existing clause/details	Modified Clause
1	II	Technical Specification – Section I	Serial No: 7	2 of 7	Fully automatic DC high voltage test kit, up to 100kV/5A Voltage Range: 0-100 kV DC Current Range: 0-5 A	Fully automatic DC high voltage test kit, up to 100kV/10mA Voltage Range: 0-100 kV DC Current Range: 0-10 mA

NOTE:- Due date of tender submission for subject package is extended up to 30.11.2021 (02.00 P.M.) and bid opening up to 30.11.2021 (04.30 P.M.)

Please ensure the submission of your offer within the due date.

This corrigendum shall be treated as part of NIT.

All the other terms and conditions of the tender enquiry remain unchanged.

All the bidders are requested to quote accordingly.

All Tender details, Corrigenda, Addenda, Amendments, Time extensions, Clarifications, etc. to the tender will be hosted on <https://eprocurebhel.co.in> , Bidders should regularly visit the website to keep themselves updated.

Thanking you,

Yours faithfully,
For and on behalf of BHEL

SWETANA SINGH

(Sr. Engineer/ Project Group-III)

Please reply to:
SWETANA SINGH (Sr. Engineer / Project Group-3)
Power Project Engineering Institute Building
HRD & ESI Complex Plot No. 25, Sector -16 A,
BHEL-PEM, Noida-201301 (U.P.)
(OFF) +91-120-4368913, (MOB.) +91 9811508427