



**BHARAT HEAVY ELECTRICALS LIMITED**  
**Heavy Power Equipment Plant, Ramachandrapuram,**  
**Hyderabad - 502 032, INDIA.**

Indent no:	
Enquiry No.:	
Due Date:	
Supplier Qtn. No.:	
Date:	

**Contact person from Purchase Dept:**

Name:  
Designation  
Phones :  
e-mail:

Purchase Dept. :

**Specification cum Compliance certificate for FOR SUPPLY OF Pressure comparison tester(pneumatic)& digital gauge:1 No.**

2. The Vendor should fill " Offered" column in compliance to specified requirement and also "Deviations" Column, where there is a deviation from the requirement. Duly filled in specification cum Compliance certificate shall be submitted along with the offer. Inadequate, Incomplete , Ambiguous, or unsustainable information against any one of the clauses of the specification/ requirement shall be treated as non-compliance. No Tick marks shall be used in the columns.

3.The offer and all documents enclosed shall be in English Language only.

**ADDRESS OF THE SUPPLIER :**

TELEPHONE NOS.:

TELEPHONE NOS.:

FAX NOS.:

FAX NOS.:

E-MAIL ADDRESS :

E-MAIL ADDRESS:

Sl. No	Description	Offered	Deviations	Remarks
	<b>Purpose :</b> The equipment is of table top type& usable to calibrate pressure gauges of range -950m bar to +30 bar. These instruments are used for checking pressure measuring instruments against master test gauges, indicators or transducers. These pneumatic test pumps shall provide a quick and easy solution for checking pressure instruments to (-)950mbar to (+)30 bar.			
2	<b>Operation :</b> Shall generates pressures to 30bar, Bench mounted design , Screw press for fine pressure adjustments.			
3	<b>connections:</b> 2 ports(1/4 and 3/8 inch NPT/BSP).			
4	<b>Operating temperature :</b> 15 to 35 °C			
5	<b>Digital gauge Accuracy:</b> 0.25%on range			
6	<b>Cost of Accessories / spares essential</b> for its operation if any shall be quoted item wise with full details of utility & functions			
7	<b>Documents &amp; manuals:</b> Shall supply two copies soft/hard of User operational manual, Service & Maintenance manuals.			
8	<b>Commissioning :</b> The supplier shall commission the equipment at BHEL, Hyderabad at free of cost.			
9	<b>Acceptance of Equipment:</b> The equipment will be accepted by BHEL upon satisfactory demonstration performance as at s.no 2 and 5 at BHEL.			
10	<b>Qualifying condition:</b> Those vendors who have Supplied and Commissioned Pressure comparison tester and digital gauge as at s.No 1 &5 and running satisfactory for at least one year. ( Full details of customer for contact shall be provided with the offer)			
11	<b>Delivery:</b> Supplier shall quote definite delivery period.			
12	<b>Packing:</b> Rigid packing to be made for all items of equipment, all accessories and other supplied items to avoid any damages/loss in transit.			