



Bid Number/बोली क्रमांक (बिड संख्या):  
GEM/2024/B/5134764  
Dated/दिनांक : 08-07-2024

### Bid Document/ बिड दस्तावेज़

| Bid Details/बिड विवरण  |   |
|--|---|
| Bid End Date/Time/बिड बंद होने की तारीख/समय  | 18-07-2024 15:00:00   |
| Bid Opening Date/Time/बिड खुलने की तारीख/समय   | 18-07-2024 15:30:00   |
| Bid Offer Validity (From End Date)/बिड पेशकश वैधता (बंद होने की तारीख से)  | 120 (Days)  |
| Ministry/State Name/मंत्रालय/राज्य का नाम  | Ministry Of Heavy Industries And Public Enterprises   |
| Department Name/विभाग का नाम   | Department Of Heavy Industry  |
| Organisation Name/संगठन का नाम   | Bharat Heavy Electricals Limited (bhel)   |
| Office Name/कार्यालय का नाम  | 10140027-hpbp Trichy  |
| Total Quantity/कुल मात्रा  | 4   |
| Item Category/मद केटेगरी   | ICU Bed (Q3)  |
| Years of Past Experience Required for same/similar service/उन्हीं/समान सेवाओं के लिए अपेक्षित विगत अनुभव के वर्ष | 3 Year (s)  |
| MSE Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से एमएसई छूट                                 | No  |
| Startup Exemption for Years of Experience and Turnover/ अनुभव के वर्षों से स्टार्टअप छूट                         | No  |
| Document required from seller/विक्रेता से मांगे गए दस्तावेज़   | Experience Criteria,Past Performance,Certificate (Requested in ATC),OEM Authorization Certificate,Additional Doc 1 (Requested in ATC),Additional Doc 2 (Requested in ATC),Additional Doc 3 (Requested in ATC)<br>*In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |
| Past Performance/विगत प्रदर्शन   | 80 %  |
| Bid to RA enabled/बिड से रिवर्स नीलामी सक्रिय किया   | Yes   |
| RA Qualification Rule  | 50% Lowest Priced Technically Qualified Bidders   |
| Type of Bid/बिड का प्रकार  | Two Packet Bid  |

**Bid Details/बिड विवरण**

|   |  |
|---|--|
| <b>Time allowed for Technical Clarifications during technical evaluation/तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय</b> | 4 Days   |
| <b>Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)</b>                                      | No   |
| <b>Payment Timelines</b>  | Payments shall be made to the Seller within <b>90</b> days of issue of consignee receipt-cum-acceptance certificate (CRAC) and on-line submission of bills (This is in supersession of 10 days time as provided in clause 12 of GeM GTC) |
| <b>Evaluation Method/मूल्यांकन पद्धति</b>   | Total value wise evaluation  |

**EMD Detail/ईएमडी विवरण**

|                   |    |
|-------------------|----|
| Required/आवश्यकता | No |
|-------------------|----|

**ePBG Detail/ईपीबीजी विवरण**

|                   |    |
|-------------------|----|
| Required/आवश्यकता | No |
|-------------------|----|

**Splitting/विभाजन**

Bid splitting not applied/बोली विभाजन लागू नहीं किया गया.

**MII Purchase Preference/एमआईआई खरीद वरीयता**

|  |    |
|--|----|
| MII Purchase Preference/एमआईआई खरीद वरीयता | No |
|--|----|

**MSE Purchase Preference/एमएसई खरीद वरीयता**

|   |     |
|---|-----|
| MSE Purchase Preference/एमएसई खरीद वरीयता | Yes |
|---|-----|

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the

Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 100% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

3. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for determining the Eligibility Criteria related to Turn Over, Past Performance and Project / Past Experience etc. This has no relevance or bearing on the price to be quoted by the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criteria in determining reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonableness and based on competitive prices received in Bid / RA process.

4. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 80% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

5. Reverse Auction would be conducted amongst first 50% of the technically qualified bidders arranged in the order of prices from lowest to highest. Number of sellers eligible for participating in RA would be rounded off to next higher integer value if number of technically qualified bidders is odd (e.g. if 7 bids are technically qualified, then RA will be conducted amongst L-1 to L-4). In case number of technically qualified bidders are 2 or 3, RA will be between all without any elimination. If Buyer has chosen to split the bid amongst N sellers, then minimum N sellers would be taken to RA round. In case Primary products of only one OEM are left in contention for participation in RA based on lowest 50% bidders qualifying for RA, the number of sellers qualifying for RA would be increased to get at least products of one more OEM (directly participated or through its reseller) if available. Further, if bid(s) of any seller(s) eligible for MSE preference is / are coming within price band of 15% of Non MSE L-1 or if bid of any seller(s) eligible for Make in India preference is / are coming within price band of 20% of non MII L-1, then such MSE / Make in India seller shall also be allowed to participate in the RA process.

## ICU Bed ( 4 pieces )

### Technical Specifications/तकनीकी विशिष्टियाँ

\* As per GeM Category Specification/जेम केटेगरी विशिष्टि के अनुसार

| Specification         | Specification Name/विशिष्टि का नाम   | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य   |
|-----------------------|--|--|
| Performance Parameter | <b>Purpose</b>   | ICU Bed are required to be used in Intensive Care Unit for comfort of Patient & facilitate comfortable transfer to & from Emergency/ OT/ Ward etc. & also to carry out point of care process including Radio-logical procedure at bed side |
|                       | <b>Type of ICU Bed Actuator</b>  | Electro-Mechanical (Motorized )  |
|                       | <b>Type of mechanism for functioning or controlling angular motion of bed part</b> | Electro-Mechanical (Motorized)   |

| Specification | Specification Name/विशिष्टि का नाम  | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य             |
|---------------|---|--|
|               | <b>Type of mechanism for functioning or controlling Height of bed</b>   | Electro-Mechanical (Motorized)   |
|               | <b>Bed top perforated</b>   | Yes  |
|               | <b>Bed should have radio translucent top (X-Ray translucent)</b>  | Yes  |
|               | <b>Type of side panel</b>   | Swing Up Down type   |
|               | <b>Number of Side panel</b>   | 4  |
|               | <b>Shape of Head &amp; Foot panel /Board</b>  | D-type   |
|               | <b>Type of Head &amp; foot Panel</b>  | Swing Up Down type   |
|               | <b>Number of hooks provided in IV rod</b>   | 4  |
|               | <b>Availability of rectangular telescopic tube box housing for tension spring</b>                             | NA for Super Smooth Crank Screw Mechanism & Hydraulic mechanism for height |
|               | <b>Degree Angle &amp; Height indicator shall be provided at both foot &amp; head side</b>                     | Yes  |
|               | <b>X-Ray film holder used from</b>  | One side   |
|               | <b>Power of motor in HP(for non motorized put NA)</b>   | 0.5 hp   |
|               | <b>Availability of patient constraint belt provided on both side of bed</b>                                   | yes  |
|               | <b>Number of caster to which braking system provided</b>  | 4  |
|               | <b>Safe working load capacity of ICU bed in kg</b>  | 250  |
|               | <b>Mattress provided with ICU bed</b>   | Yes  |
|               | <b>ICU mattress must X-ray translucent</b>  | Yes  |
|               | <b>Mattress shall be made of High resilient &amp; bio-density foam</b>  | Yes  |
|               | <b>Mattress should be translucent to allow radiography using portable X-ray machines</b>                      | Yes  |
|               | <b>Mattress should be made in cube cut design &amp; independent cubes promote air flow to reduce moisture</b> | Yes  |

| Specification                    | Specification Name/विशिष्टि का नाम   | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|----------------------------------|--|--|
|                                  | <b>Material for ICU Bed Mattress</b>   | PU Foam  |
|                                  | <b>Therapeutic Weight limit for mattress in kg</b>   | 175  |
|                                  | <b>Power Supply ( as appropriate fitted with Indian plug)</b>  | Single phase 230 Volt, 50 Hz ( AC Supply )                     |
|                                  | <b>Facility of Resettable overcurrent breaker shall be fitted for protection</b>   | Yes  |
|                                  | <b>Warranty (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)</b> | 2  |
| Standard Parameter               | <b>Conformity to standard for safety &amp; electromagnetic compatibility</b>   | IEC-60601-1-2:2001 Standard or IS-13450                        |
| Dimensional & Material Parameter | <b>Material for the frame of bed</b>   | MS   |
|                                  | <b>Material for side railing/ Side Safety guard</b>  | ABS Plastic  |
|                                  | <b>Material for Head &amp; Foot Panel/Board</b>  | ABS Plastic  |
|                                  | <b>Material for Bed top Section</b>  | ABS Plastic  |
|                                  | <b>Material of wheels</b>  | Polyester  |
|                                  | Material for IV rod with chromium plated   | SS   |
|                                  | <b>Maximum Adjustable Back Rest Angle in Degree</b>  | 0-70   |
|                                  | <b>Maximum Adjustable Knee Rest Angle in Degree</b>  | 0-40   |
|                                  | <b>Maximum Trendlenburg Angle in Degree with tolerance <math>\pm 2</math> degree</b>   | 15   |
|                                  | <b>Maximum Reverse Trendlenburg Angle in Degree with tolerance <math>\pm 2</math> degree</b>   | 15   |

| Specification         | Specification Name/विशिष्टि का नाम  | Bid Requirement/बिड के लिए आवश्यक (Allowed Values)/अनुमत मूल्य |
|-----------------------|---|--|
|                       | Mean coil diameter of tension spring used in counter weight mechanism in mm | 40   |
|                       | Spring coil diameter used in counter weight mechanism in mm                 | 6  |
|                       | Clearance between Bed Base frame and Floor surface in mm                    | 150  |
|                       | Length of bed in mm with 2% tolerance                                       | 2100-2200  |
|                       | Width of bed in mm with 2% tolerance  | 1000-1050  |
| Additional Parameters | Spares ICU Bed Mainframe supplied   | Not supplied   |
|                       | Spares Pair of Bed Ends, detachable supplied                                | Not supplied   |
|                       | Spares Collapsible side rails supplied                                      | Not supplied   |
|                       | Spares IV Rods supplied   | 1  |
|                       | Spares Mattress simmilar to one supplied with ICU bed                       | not supplied   |
| Reports               | Electric Shock Protection level-Class-B/class-1                             | Yes  |

#### Additional Specification Parameters - ICU Bed ( 4 pieces )

| Specification Parameter Name                    | Bid Requirement (Allowed Values)  |
|---|---|
| Hand set  | Bed should be supplied with a hand set to operate   |
| service centre                                  | The availability of service centre should be there in Tamilnadu   |
| test Reports/certificate                        | Copies of test Reports/certificates and other certifications to be furnished to buyer on demand at time of supply |
| Electric Shock Protection level-Class-B/class-1 | Electric Shock Protection level-Class-B/class-1   |
| Certifications for Product                      | European CE   |

\* Bidders offering must also comply with the additional specification parameters mentioned above.

**Consignees/Reporting Officer/परेषिती/रिपोर्टिंग अधिकारी and/ तथा Quantity/मात्र**

| S.No./क्र. सं. | Consignee Reporting/Officer/परेषिती/रिपोर्टिंग अधिकारी | Address/पता   | Quantity/मात्र | Delivery Days/डिलीवरी के दिन |
|----------------|--|---|----------------|------------------------------|
| 1              | Vijayakumar Munusamy                                   | 620014,MATERIALS MANAGEMENT, IV FLOOR, BUILDING -24, HIGH PRESSURE BOILER PLANT, BHARAT HEAVY ELECTRICALS LIMITED, TIRUCHIRAPPALLI 14 | 4              | 28                           |

**Special terms and conditions-Version:4 effective from 09-01-2024 for category ICU Bed**

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules made there under as amended till date will always be applicable. This will include all notifications issued by *Central Drugs Standard Control Organisation(CDSCO)*, Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
  2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
  3. The price offered by the seller shall not, in any case exceed the DPCO controlled price, if any, fixed by the Central/State Government, the Maximum Retail Price (MRP) and the selling price. The seller must reduce the prices if there is any reduction in DPCO ceiling price, if any.
  4. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
  5. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address,telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled.Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.
  6. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
  7. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
  8. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee

before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.

OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.

9. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
10. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
11. **Software:** All software updates should be provided free of cost during warranty period.
12. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.f

## **Buyer Added Bid Specific Terms and Conditions/क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें**

### **1. Generic**

Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address

Manager / Stores - 12  
Fossil Boilers  
BHEL Boiler Plant Project P.O.  
Tiruchirappalli - 620 014  
TAMILNADU, INDIA

### **2. Generic**

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.

### **3. Generic**

Bidder shall submit the following documents along with their bid for Vendor Code Creation:

- a. Copy of PAN Card.
- b. Copy of GSTIN.
- c. Copy of Cancelled Cheque.
- d. Copy of EFT Mandate duly certified by Bank.

### **4. Generic**

Supplier shall ensure that the Invoice is raised in the name of Consignee with GSTIN of Consignee only.

## 5. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of GST portal confirming payment of GST.

## 6. **Purchase Preference (Centre)**

Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail the Purchase preference, the bidder must be the manufacturer of the offered product in case of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service. If L-1 is not an MSE and MSE Seller (s) has/have quoted price within L-1+ 15% of margin of purchase preference /price band defined in relevant policy, such Seller shall be given opportunity to match L-1 price and contract will be awarded for percentage of 100% of total value.

## 7. **Certificates**

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.

## 8. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

## **Disclaimer/अस्वीकरण**

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the

case may be.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

[This Bid is also governed by the General Terms and Conditions/ यह बिड सामान्य शर्तों के अंतर्गत भी शासित है](#)

In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws./जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा।

**---Thank You/धन्यवाद---**