

SPECIFICATION FOR DIGITAL HAND HELD DP INDICATOR

General: Handheld battery operated digital pressure manometer for measuring differential pressure (of air flow pitot tube) with an accuracy of ± 0.25 % F.S. with bold LCD display of at least 1 cm high.

Technical Specification:

Pressure range:	-100 to 100 mbar
Resolution:	0.1 mbar
Accuracy:	0.25% or better
Zero Function:	Easy accessible zero adjustment knob
Display	Bold LCD Digital around ½"
Indication:	Low battery
Process connection:	Quick fit connector with suitable dual tubing of 2 metre length to suit pitot tube pressure connector.

Accessory:

- Ellipsoidal nosed pitot tube of 1 metre & 1.5 metre length with k factor of 1 with calibration certificate.
- Carrying case
- Operating manual
- Factory calibration certificate
- Batteries for instrument initial operation.

PLEASE NOTE:

1. The instrument offered can be better than the above specifications
2. Should have service facility in India with factory trained engineers to take care of the equipment supplied and to reduce the down time.
3. The product catalogue should contain only the product offered in depth preferably highlighting the model offered.
4. Two sets of original product catalogue to be enclosed with the offer, without this offer will be rejected totally.
5. Vendor should submit the offer with detailed specification of the product offered. The offers given as **“as per your specifications”** will be rejected.