

**SPECIFICATION FOR EARTH HI TESTER**

Ground Resistance:	Up to 40Ω with an accuracy of ±10% or better 40Ω up to 500Ω with an accuracy of ±5% or better
Ground Voltage:	0 to 10 Volt with an accuracy of ±5% or better
Principle of operation:	Voltage Comparison method
Measuring Frequency:	100 KHz to 1.5MHz approx
Battery:	Easily available Alkaline battery With auto power off function and low battery indication
Operating duration:	8 hours continuous with above battery
Display:	Bold LCD with 4 digits
Short circuit current:	around 3mA

Accessory:

1. 3nos Test leads of 3 different colours with cable winder
2. Carrying case
3. Operation manual
4. Set of batteries for initial operation
5. Factory test certificate

**Please Note:**

1. The instrument offered can be better than the above specifications
2. Should have service facility in India preferably with factory trained engineers to take care of the equipment supplied and to reduce the down time.
3. The product catalogue should contain only the product offered in depth preferably highlighting the model offered.
4. Two sets of original product catalogue to be enclosed with the offer, without this offer will be rejected totally.
5. Vendor should submit the offer with detailed specification of the product offered. The offers given as “**as per your specifications**” will be rejected