

Special Remarks for enquiry no.E7110462

1. Calibration certificate traceable to National/International standard to be provided.
2. The instrument along with all necessary accessories should fulfill requirement of standards, IEC: 60270 & is: 6209.
3. Similar instrument must be in operational condition for more than three years in India & performance certificate of the above instrument obtained from the utilizer's end (customer) should be submitted with the technical bid.
4. Offer of recommended spares for five years for trouble free operation to be provided along with instrument, should be submitted with the technical bid.
5. Instrument should have adequate and reliable safety interlocks / devices to avoid damage to system, work piece and the personnels.
6. If any safety/environmental protection is required, it should be provided in the instrument by vendor.
7. Commissioning spares required for commissioning of the instrument should be available at the time of commissioning at our works.
8. BHEL testing engineer and maintenance person should be trained at BHEL works in the area of operation and maintenance for instrument.
9. Inspection at supplier works by BHEL before dispatch.
10. Operation & maintenance manual with details circuit diagrams, schematics should be supplied with the consignment.