



PQC for PR-0149284573 dated 26.08.2026**Annexure-I : Pre Qualifying Criteria**

Sl.No.	Description of Pre Qualification Requirement	Bidder Response
1	The Vendor should be M/S. TEC-SOL INDIA PRIVATE LIMITED.	



P. VENKATACHALAMOORTHY
ENGINEER
PLANT SERVICES / SSTP
BHEL, TRICHY - 620 014

TDC for PR-0149284573 dated 26.08.2026		
	DESCRIPTION	Vendor Confirmation
PR Item No		
10	Application :	40 Ton UTM – Hydraulic Pressure Measurement,
	Sensor Type	Industrial Pressure Transducer,
	Pressure Range	0–250 Bar,
	Output :	4–20 mA DCM,
	Supply Voltage:	5 - 10 V DC,
	Accuracy:	±0.5% F.S. or better,
	Pressure Port:	1/4" BSP Male,
	Wetted Material:	Stainless Steel,
	Electrical Connection:	DIN / M12 Connector,
	Protection:	IP65 / IP67,
	Operating Temperature :	-20°C to +80°C,
	Over Pressure :	≥150% F.S.,
	Burst Pressure :	≥200% F.S.,
	Response Time :	≤5 ms,
	Calibration:	Calibration with traceable certificate,
	Accessories:	Mounting adaptor, connector and required fittings,
	Scope:	Supply, Installation & Calibration,


P. VENKATACHALAMOORTH
ENGINEER
PLANT SERVICES / SSTP
BHEL, TRICHY - 620 014