

9

**TECHNICAL SPECIFICATION FOR ENT TREATMENT UNIT  
FOR YEAR 2011-2012**


Basic unit for OPD for diagnostic and Therapy for Ear, Nose, Throat. It can help to diagnose with optimal efficiency and easy to work without mobilizing the entire unit.

**TECHNICAL SPECIFICATION**

1. Spraying Device with a built in automatic micro switch & spray directly coupling with compressor motor for Nasal Spray & Laryngeal Spray,
2. Suction device is operated by an automatic switch system
3. Examination Lamp & arm set to be assembled on the post pole
4. Instrument Tray which can pull out for checking inside of unit
5. can with cap for Swap & Medicine
6. Waste container for instrument & waste materials
7. Medicine bottle
8. Mirror Warmer
9. Endoscope Light Source
10. Head light with adjustable head band  
(Halogen)with suitable fibro-optic cable
11. Doctor's Chair Hydraulic
12. X-Ray view
13. Head Mirror
14. Tray 2.
15. Endoscope Tray.
16. Monitor Medical Grade.
17. OPD Instruments 20.
18. Endoscope Camera System.
19. Sinuscope 0°.
20. CCD camera with approximately 400,000pixel integrated light source with fibro-optic cable.
21. 14" TFT Monitor


**Term & conditions**

1. Warranty for the entire system for 2 years from date of commissioning
2. Annual maintenance to be quoted for 5 years after the warranty period
3. Service manual
4. The system will be declared commissioned only after successful commissioning of the unit.

  
22/2/2012  
Dr.S.P.Singh  
Sr. Specialist (ENT)

Pre-Qualifying Conditions for Tender No.- C/7029/11/4339T

1. The Manufacturer should have standing of manufacturing the equipment for minimum 5 years.
2. The Manufacturer should furnish the list of customers where same model of equipment has been commissioned in last 5 years from the date of opening of the Tender.
3. The vendor should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactorily for last at least 1 year from the date of opening of the Tender. Also furnish the details of customer/company where the equipment has been installed with contact address and telephone no. of the customer.
4. The vendor should ensure regular after sales service including spares for at least five years after expiring of Warranty period.



Dr. S.K. Gupta  
AGM (Surgery)

27/3/12