

9

**TECHNICAL SPECIFICATION FOR ENT TREATMENT UNIT
FOR YEAR 2011-2012**


Basic unit for OPD for diagnostic and Therapy for Ear, Nose, Throat. It can help to diagnose with optimal efficiency and easy to work without mobilizing the entire unit.

TECHNICAL SPECIFICATION

1. Spraying Device with a built in automatic micro switch & spray directly coupling with compressor motor for Nasal Spray & Laryngeal Spray,
2. Suction device is operated by an automatic switch system
3. Examination Lamp & arm set to be assembled on the post pole
4. Instrument Tray which can pull out for checking inside of unit
5. can with cap for Swap & Medicine
6. Waste container for instrument & waste materials
7. Medicine bottle
8. Mirror Warmer
9. Endoscope Light Source
10. Head light with adjustable head band
(Halogen)with suitable fibro-optic cable
11. Doctor's Chair Hydraulic
12. X-Ray view
13. Head Mirror
14. Tray 2.
15. Endoscope Tray.
16. Monitor Medical Grade.
17. OPD Instruments 20.
18. Endoscope Camera System.
19. Sinuscope 0°.
20. CCD camera with approximately 400,000pixel integrated light source with fibro-optic cable.
21. 14" TFT Monitor


Term & conditions

1. Warranty for the entire system for 2 years from date of commissioning
2. Annual maintenance to be quoted for 5 years after the warranty period
3. Service manual
4. The system will be declared commissioned only after successful commissioning of the unit.


23/2/2012
Dr.S.P.Singh
Sr. Specialist (ENT)

Pre-Qualifying Condition for Tender No.- C/7029/11/4339T

Vendor should furnish performance certificate from at least one customer where ENT Therephy Unit has been installed and has been working satisfactorily. Contact details of the customer should be submitted.


15/11/22
Dr. S.K. Gupta
AGM (Surgery)