

Technical specification of lab centrifuge

1. Centrifuges to separate various components of Blood and any other liquid sample for analysis
2. Table top version
3. Tube Capacity :No. 16 – 24 :Size 15 ml
4. Should have a digital timer
5. Should have a digital speed meter
6. Centrifuge complete with Swing out head and basic rotors
7. Speed regulators for speed 1000 rpm to 4000 rpm
8. Maintenance-free brushless drive motor with exact speed preselection and display.
9. Speed indicator with timer
10. Power input to be 220-240VAC, 50Hz as appropriate fitted with Indian plug
11. User manual in English
12. Certificate of calibration and inspection.

29/11/11

Case A
Hear

Dr Anjali Mishra
Dy Mgr. Pathology

SE

Pre-Qualifying Conditions for Tender No.-C/7029/11/4332T

1. The Manufacturer should have standing of manufacturing the equipment for minimum 5 years.
2. The Manufacturer should furnish the list of customers where same model of equipment has been commissioned in last 5 years from the date of opening of the Tender.
3. The vendor should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactorily for last at least 1 year from the date of opening of the Tender. Also furnish the details of customer/company where the equipment has been installed with contact address and telephone no. of the customer.
4. The vendor should ensure regular after sales service including spares for at least five years after expiring of Warranty period / *guarantee period*

Sm

Dr. S.K. Gupta
27/3/12

Dr. S.K. Gupta
AGM (Surgery)