

Bharat Heavy Electricals Limited, Jhansi

Hospital Administration



Specification Rev 01 of Road Ambulances to Indent No 679590030 dt 06/09/2019

63

Basic Life Support Road Ambulance: Ambulance ergonomically designed, suitably equipped and appropriately staffed for the transport and treatment of patients requiring non-invasive airway management / basic monitoring. Equipment and Accessories required for AIS 125, Type 'C' with at least one Auto Loading Stretcher.

Essential Requirement of Ambulance		
Basic Vehicle:		
Particulars	Bidder to Tick '✓' is whichever applicable	Document enclosed
Engine: Petrol/ Diesel : Min 1150CC		
Size Minimum (L*W*H): 3650*1450*1600		
Air Conditioner		
Fabrication/ Interior/ Exterior:		
Homogeneous non static, mark resistant, scratch proof, anti-bacterial, Anti-Skid Vinyl flooring material with water proof fire retardant ply-board.		
Under floor, Cabinet for First-aid medicine and medical consumable .		
AIS Approved Auto-Loading stretcher for patient transportation in MS with ramp.		
B' type cylinder-10 liter water capacity lying under the Flooring in MS bracket with pressure gauge, Humidifier and mask		
Hooks for IV bottle's- 2No with IV holding attachment		
Stainless Steel Wash Basin with Water		

Rev 01: Rev 01 is being issued to modify Para 13 as written in bold letters above.

We agree that our bid is liable to be rejected in case (a) the relevant information is not filled, (b) documentation is incomplete, (c) there are false declarations, and (d) if we do not meet the mandatory criteria on Financial Turnover, Experience & Mandatory Resource requirements.

सामग्री प्रबन्धन

15 NOV 2019

भेल, झाँसी

(Signature of the Supplier)

With Name, Designation & Seal)

Page 1/4

(ग्यानendra सिंह)
(GYANENDRA SINGH)
वरिष्ठ प्रबन्धक (अस्पताल प्रशासक)
Sr. Manager (Hospital Adm.)
बी.एच.ई.एल., झाँसी
B.H.E.L., JHANSI

Bhawan
13/11/2019
(डा. 0 प्रमुख (सहस्र))

(डा. 0 सुजीत सिंह)
(DR. S. SINGH)
चिकित्सा अधीक्षक
Medical Superintendent
भेल, झाँसी
BHEL Hospital, JHANSI

Bharat Heavy Electricals Limited, Jhansi

Hospital Administration



Specification Rev 01 of Road Ambulances to Indent No 679590030 dt 06/09/2019

62


Desirable Requirement of Ambulance


	Bidder to Tick '✓' is whichever applicable	Document enclosed
1. Bedding equipment 2. Material for treatment of wounds-First-aid box 3. Seat belt cutter 4. Stainless steel waste bin. 5. Needle cutter (Manually operated) 6. GPS 7. Portable Cooling Box 8. Warning Triangle 9. Fire Extinguisher-ABC type 1 kg 10. Public address system with light bar-LED with musical siren 11. External Graphics as per AIS Norms and BHEL Logo, 12. Manually Operated Suction Machine 13. Equipment and Accessories required for AIS 125 1Auto Loading Stretcher • <i>The Ambulance Cot Including all accessories should be EN 1865 Certified.</i> <u>OR the Cot should be supplied with an EN 1789 certified fixation system.</u> • Made of Stainless steel/MS with powder coated • The stretcher should Load seamlessly and		

Rev 01: Rev 01 is being issued to modify Para 13 as written in bold letters above.

We agree that our bid is liable to be rejected in case (a) the relevant information is not filled, (b) documentation is incomplete, (c) there are false declarations, and (d) if we do not meet the mandatory criteria on Financial Turnover, Experience & Mandatory Resource requirements.

Page 2/4


(Gyanendra Singh)
 वरिष्ठ प्रबन्धक (अस्पताल प्रशास) Sr. Manager (Hospital Adm.)
 बी.एच.ई.एल., झांसी B.H.E.L., Jhansi


(Dr. H. J. Singh)
 चिकित्सा अधीक्षक Medical Superintendent
 बी.एच.ई.एल., झांसी BHEL Hospital Jhansi

(Dr. H. J. Singh)
 चिकित्सा अधीक्षक Medical Superintendent
 बी.एच.ई.एल., झांसी BHEL Hospital Jhansi

(Signature of the Supplier)

With Name, Designation & Seal)

no manual intervention vis a vis the locking mechanism, wheels, etc. should be required after loading in the ambulance to close the rear doors. • The stretcher mattress should be waterproof and upholstered with fireproof material. 14.2.Scoop Stretcher • Lightweight , Portable Aluminum Scoop Stretcher • commonly used for Spinal bone fractured patients • Folded into two parts • Adjustable length • (L X W X H) 164 x 43 X 8 cm (Min.) • (L X W X H) 201 x 43 X 8 cm (Max.)		
--	--	--

Rev 01: Rev 01 is being issued to modify Para 13 as written in bold letters above.

We agree that our bid is liable to be rejected in case (a) the relevant information is not filled, (b) documentation is incomplete, (c) there are false declarations, and (d) if we do not meet the mandatory criteria on Financial Turnover, Experience & Mandatory Resource requirements.

(Signature of the Supplier

With Name. Designation & Seal)

Page 3/4

(ग्यानेन्द्र सिंह)
(GYANENDRA SINGH)
वरिष्ठ प्रबन्धक (अस्पताल प्रशा.)
Sr. Manager (Hospital Adm.)
बी.एच.ई.एल., झांसी
B.H.E.L., JHANSI

Bharat Heavy Electricals Limited
13/11/2019
(डॉ. सुजीत सिंह)
(DR. S. J. SINGH)
चिकित्सा अधीक्षक
Medical Superintendent
मेल. हॉस्पिटल, झांसी
BHEL Hospital, JHANSI

Bharat Heavy Electricals Limited, Jhansi

Hospital Administration



Specification Rev 01 of Road Ambulances to Indent No 679590030 dt 06/09/2019

15. Internal lights (LED) 16. AC Socket- for all medical /electrical equipment. 17. LED Emergency light bar mounted on the front top of driver cabin. 18. Fans DC 8" -in patient cabin and 1 in driver cabin. 19. Electric Protection Circuit with heavy duty 12 V DC Cut off switch located near driver seat for easy accessibility. 20. Light Bars at Vehicle Roof and Warning lights at both sides		
---	--	--

Rev 01: Rev 01 is being issued to modify Para 13 as written in bold letters above.

We agree that our bid is liable to be rejected in case (a) the relevant information is not filled, (b) documentation is incomplete, (c) there are false declarations, and (d) if we do not meet the mandatory criteria on Financial Turnover, Experience & Mandatory Resource requirements.

(Signature of the Supplier

With Name. Designation & Seal)

Page 4/4

(ग्यानendra सिंह)
(GYANENDRA SINGH)
वरिष्ठ प्रबंधक (अस्पताल प्रशास.)
Sr. Manager (Hospital Adm.)
बी.एच.ई.एल., झांसी
BHEL, JHANSI

Bharat Heavy Electricals Limited
(BHEL) Jhansi
प्रमुख (सिजी)

(डॉ. अजीत सिंह)
Dr. A. J. SINGH
निदेशक अधीक्षक
Medical Superintendent
बेल हॉस्पिटल, झांसी
BHEL HOSPITAL, JHANSI