

Tender Notice No. : **TTX/IT/CAL/03**

Dated: 25.09.2019

Scope of Work and Payment

| S. No. | Scope | Confirmation by Calibration Agency |
|---------------|---|---|
| 1 | Calibration of the following equipment: “Standard VT, Electronic Potential divider (EPD), A(I/V)TTS (Automatic instrument/voltage transformer test system) & Burdens as per Annexure-3” Quantity : 11 Nos. | Complied/ Not Complied |
| 2 | The calibration work will have to be completed within 30 days from date of P.O. placement at BHEL Bhopal works. Delay in execution of work may cause penalty of 0.5% of contract value per week upto maximum of 10%. | Complied/ Not Complied |
| 3 | All the ranges shall be covered during calibration. Please refer the Calibration accuracy details enclosed in Annexure-3. Should there be any comments or range which are technically not feasible; the same may please be brought out. | Complied/ Not Complied |
| 4 | The instrument will be calibrated at the works of BHEL Bhopal (Instrument Test Area,Block-3). All items required to calibrate items shall be brought by the Calibration Agency. Any facility or necessary arrangement required for calibration of Items shall be clearly written in submitted offer. | Complied/ Not Complied |
| 5 | The Standard Equipment/ Instruments used for calibration shall have capability (class/accuracy/stability) to maintain proper accuracy during calibration and traceability to National/ International Standards. Calibration agency to specify in their offer. Documents for traceability of standard and instrument used for calibration shall be submitted with offer. | Complied/ Not Complied |
| 6 | Calibration stickers bearing equipment details, calibration date, calibration status and next due date shall be put on the Instrument after calibration. | Complied/ Not Complied |
| 7 | Detail calibration certificate of the instrument calibrated shall be prepared in line with ISO 17025:2005/2017 and should be submitted within 7 days from date of completion of calibration. There shall be separate certificate for each 11 instruments calibrated. | Complied/ Not Complied |
| 8 | The payment shall be made after completion of work and submission of Final Invoice/ Bill. | Agreed/ Not Agreed |
| 9 | Deviation, if any, shall be clearly brought out. | Complied/ Not Complied |
| 10 | The Party shall confirm all the above points failing which their offer may not be considered. | Complied/ Not Complied |