

ANNEXURE-A:

- a) Test Voltages: 250 V, 500 V, 1000V, 2500V, 5000V.
- b) Insulation Resistance Range: 10K Ω to 5 T Ω .
- c) Accuracy: $\pm 5\%$ @ 1 T Ω .
- d) Automatic Discharge of potential.
- e) Internal Rechargeable Battery
- f) Vendor has to supply the Insulation Tester Battery in extra as a mandatory spare.
- g) Guarantee of 1 year from the date of receipt and acceptance. PBG of 10 % of PO Value to be retained for 02 months beyond guarantee period. For any manufacturing defect/Breakdown of the equipment, Supplier has to attend the problem within a duration of 10 days of intimation of the same by BHEL
- h) Test & Calibration Certificates of the Instrument from the NABL accredited lab

Ravindra
11/01/17
रविन्द्र कुमार
वरि. अभियंता (गुणता)
बी.एच.ई.एल., रुद्रपुर
उत्तराखण्ड - 263153

RAVINDRA KUMAR
Sr. Engineer (Q.C)
B.H.E.L., Rudrapur
Uttarakhand-263153