

**Annexure - 1****TECHNICAL SPECIFICATION OF INDENT No. - 631300002****Item No. - 1 (Density Strip)**

|   |                                  |                                                                                                                                               |
|---|----------------------------------|-----------------------------------------------------------------------------------------------------------------------------------------------|
| 1 | <b>Make :</b>                    | Shall be made of AGFA D7 or equivalent film with proper identification no. inscribed on step tablet.                                          |
| 2 | <b>Range :</b>                   | Optical densities ranging from less than 0.3 to greater than 4.0.<br><br>Shall have 15 to 17 steps with approximate intervals of 0.3 or less. |
| 3 | <b>Shelf life :</b>              | Shall be at least 4 years from the date of supply.                                                                                            |
| 4 | <b>Calibration Certificate :</b> | Calibration certificate (issued by nationally or internationally accredited laboratory) shall be provided along with the density strip.       |
| 5 | <b>Packing :</b>                 | Shall be packed in Lightproof, Moisture proof and Damage proof packing.                                                                       |

*Bhupendra*  
12.06.2020

*R.K. Jindal*  
12.06.2020

*V.K. Dongre*  
12.06.2020  
V.K. DONGRE  
Addl. General Manager  
Quality Control (QND)  
B.H.E.L., BHOPAL