

 BPC 0007	भारत हेवी इलेक्ट्रिकल्स लिमिटेड, पिपलानी, भोपाल - ४६२०२२(भारत) सामग्री प्रबंधन विभाग BHARAT HEAVY ELECTRICALS LIMITED, PIPLANI, BHOPAL – 462022(INDIA) MATERIALS MANAGEMENT DIVISION	ENQUIRY NO E3903017 ENQUIRY DATE 16/12/20 ENQUIRY DUE DATE 06/01/21						
ENQUIRY TWO PART BID	<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td style="width:33%;">TIN NO. 23573600001</td> <td style="width:33%;">ECC NO. AAACB4146PXM009</td> <td style="width:33%;">MPCT NO. HEL/05/01/0001/S15/11/79</td> </tr> <tr> <td>PHONE: 91-755-2500100 (7 LINE)</td> <td>FAX: 91-755-2500023</td> <td>WWW.BHEL.BHOPAL.COM</td> </tr> </table>		TIN NO. 23573600001	ECC NO. AAACB4146PXM009	MPCT NO. HEL/05/01/0001/S15/11/79	PHONE: 91-755-2500100 (7 LINE)	FAX: 91-755-2500023	WWW.BHEL.BHOPAL.COM
TIN NO. 23573600001	ECC NO. AAACB4146PXM009	MPCT NO. HEL/05/01/0001/S15/11/79						
PHONE: 91-755-2500100 (7 LINE)	FAX: 91-755-2500023	WWW.BHEL.BHOPAL.COM						

SUPP NAME AND ADDRESS	SUPP CODE	REV CD	REV NO	REV DATE	NO OF CATY2	NO OF CATY3	ENQ NO OF ITEMS	INDENT NO
OFFICE COPY		1	0	NA	0	0	4	638400011
	GUARANTEE CERTIFICATE		Y	WELL PACKED DELIVERY REQD. AT BHEL BHOPA L				
	TEST CERTIFICATE		Y					
	INSTRUCTION BOOKLET		N	TECHNICAL CONDITION GAS DETECTION TUBES WITH STROKE PUMP FOR MEDICAL O XYGEN TESTING				
	SAMPLE		N					
	GATE PASS		Y	INSPECTION CONDITION AT BHEL BHOPAL				

NOTE: QUOTE PRICE BOTH IN FIGURES & WORDS. IN CASE OF MISMATCH PRICE IN WORDS WILL BE VALID, QUOTATIONS NOT BEARING ENQUIRY NO AND DUE DATE LIABLE TO BE REJECTED.

SL NO	MATERIAL CODE	DESC	UNIT	ITEM QTY	LOT NO	LOT QTY	DEST	DELIVERY DATE
1	NP3840210026	CARBON MONOXIDE GAS DETECTION TUBE, PACK OF 10 TUBES, AS PER INDIAN PHARMACOEPIA ADDENDUM 2005, CARBON MONOXIDE DETECTION IN MEDICAL OXYGEN GAS USING STROKE PUMP, LOWER RANGE: LESS THAN OR EQUAL TO 2.5 PPM V/V, UPPER RANGE: MORE THAN OR EQUAL TO 10 PPM V/V.	PK	45.000	1	20.000	384	20/01/21
					2	25.000	384	20/05/21
2	NP3840210027	CARBON DIOXIDE GAS DETECTION TUBE, PACK OF 10 TUBES, AS PER INDIAN PHARMACOEPIA ADDENDUM 2005, CARBON DIOXIDE DETECTION IN MEDICAL OXYGEN GAS USING STROKE PUMP, LOWER RANGE: LESS THAN OR EQUAL TO 100 PPM V/V UPPER RANGE: MORE THAN OR EQUAL TO 500 PPM V/V.	PK	45.000	1	20.000	384	20/01/21
					2	25.000	384	20/05/21
3	NP3840210028	WATER VAPOUR DETECTION TUBE, PACK OF 10 TUBES, AS PER INDIAN PHARMACOEPIA ADDENDUM 2005, WATER VAPOUR DETECTION IN MEDICAL OXYGEN GAS USING STROKE PUMP, LOWER RANGE: LESS THAN OR EQUAL TO 65 PPM V/V, UPPER RANGE: MORE THAN OR EQUAL TO 200 PPM V/V.	PK	45.000	1	20.000	384	20/01/21
					2	25.000	384	20/05/21
4	NP3840210029	STROKE PUMP AS PER INDIAN PHARMACOEPIA ADDENDUM 2005, FOR STROKES OF GAS DETECTION TUBES FOR MEDICAL OXYGEN. TESTING. COMPATIBLE WITH GAS DETECTION TUBES.	NO	1.000	1	1.000	384	20/01/21

REMARK (1) OFFER TO BE SUBMITTED ONLINE ONLY THROUGH EPROCUREBHEL.CO.IN (2) QUOTATION VALIDITY SHOULD BE 90 DAYS. (3) ALL TERMS & CONDITIONS OF TENDER AS PER ANNEXURE - A AND ANNEXURE-BP200102: ENQUIRY- GTC. (4) PRE QUALIFICATION CRITERIA AS PER ENCLOSED PQC. SUBMIT THE DOCUMENTS REQUIRED AS PER PQC. (5) ACCEPT ALL TERMS & CONDITION OF BHEL AND SUBMIT ANNEXURE-A & ANNEXURE BP200102 AFTER FILL, SIGN & SEAL ON EACH PAGE OF THESE DOCUMENTS.

DRAWING N	PURCH SPEC Y	CATALOUGE N	PLAN N	TWO PART BID Y
--------------	-----------------	----------------	-----------	-------------------

NOTE:BHEL,BHOPAL'S Standard Terms & Conditions BP4255/MM5533 rev02 for imports and BP4255/MM5527 rev03 for indigeneous procurements form a part of this Enquiry. Bidders may obtain from us copies of these terms and conditions if not already available.

SPECIAL REMARKS	NAME : KAILASH CHANDRA MEENA DESG : SR.ENGINEER SIGN & SEAL
------------------------	--

E
N
Q
U
I
R
Y



BPC 0007

**ENQUIRY
TWO PART BID**

भारत हेवी इलेक्ट्रिकल्स लिमिटेड, पिपलानी, भोपाल - ४६२०२२(भारत)
सामग्री प्रबंधन विभाग
BHARAT HEAVY ELECTRICALS LIMITED, PIPLANI, BHOPAL – 462022(INDIA)
MATERIALS MANAGEMENT DIVISION

TIN NO. 235736000001	ECC NO. AAACB4146PXM009	MPCT NO. HEL/05/01/0001/S15/11/79
----------------------	-------------------------	-----------------------------------

PHONE: 91-755-2500100 (7 LINE) FAX: 91-755-2500023 WWW.BHEL.BHOPAL.COM

ENQUIRY NO E3903017
ENQUIRY DATE 16/12/20
ENQUIRY DUE DATE 06/01/21

**E
N
Q
U
I
R
Y**

S.NO	INDENT NO	ITEM NO	CATEGORY	ENQUIRY QTY.
1	638400011	1	788500	45.000
2	638400011	2	788500	45.000
3	638400011	3	788500	45.000
4	638400011	4	788500	1.000

NOTE:BHEL,BHOPAL'S Standard Terms & Conditions BP4255/MM5533 rev02 for imports and BP4255/MM5527 rev03 for indigeneous procurements form a part of this Enquiry.Bidders may obtain from us copies of these terms and conditions if not already available.

SPECIAL REMARKS

NAME : KAILASH CHANDRA MEENA
DESG : SR.ENGINEER

SIGN & SEAL