

BHARAT HEAVY ELECTRICALS LTD.

SPECIFICATION No. :
ITEM No. :
LOCATION : KASTURBA HOSPITAL(OPERATION THEATER)

SPECIFICATION FOR SPINE SURGERY

1.0 Requirement / Function :- Operating Facility for spine surgery.

2.0 Specification / Features:-

Spinal frame :

- 2.1** Top with rubber mattress.
- 2.2** Adjustable width as per requirement.
- 2.3** Can be used in all operation Table.
- 2.4** Anti corrosive stainless steel material.

3.0 Scope of Supply-

- 3.1** Equipment Meeting our above specification.
- 3.2** Installation, commissioning, Training & proving of equipment by supplier.
- 3.3** Two sets of operation & maintenance manuals with circuit diagrams.
- 4.0** Quoted equip. should be proven eqpt. And suppliers should furnish list of customers where quoted eqpt. Has been supplied earlier with necessary certificates for good performance and after sales service.
- 5.0** Eqpt. Should be suitable to operate on BHEL Power supply of 230 V+- 15% 50Hz+ 3% single phase AC supply and suitable voltage stabilize/CVT if recommended by the supplier should be supplied with equipment.
- 6.0** Supplier must quote all material/facility/service required for satisfactory installation and commissioning of equip. BHEL scope if any to be mentioned specifically. Anything not mentioned specifically but required for satisfactory commissioning/proving of Eqpt. Will have to be arranged by supplier.
- 7.0** Bidders must furnish point to point replies against all the points of our above specification failing which their offer may not be considered.