

BHARAT HEAVY ELECTRICALS LIMITED, BHOPAL

SPECIFICATION NO. : **I**
ITEM NO. : **II**
LOCATION : NEURO, OT KASTURBA HOSPITAL
SPECIFICATIN FOR : 1 QTY

1. REQUIREMENT / FUNCTION **Operating facility for Brain Tumour**

2. Specification/Features

- 2.1 Four pin Fixation Neurosurgery skull clamp for micro Neurosurgery
- 2.2 Semi circular Based frame
- 2.3 Self retaining retractors
- 2.4 Head holder with frame
- 2.5 Can be applied to any OT table

3. Scope of Supply

No consumables other optional accessories required are.

- 3.1 Horse shoe with pulley traction
- 3.2 Sitting crossbar (adjustable)

- 4. Quoted eqpt.should be proven and suppliers should furnish list of customers where quoted Eqpt. Has been supplied earlier with necessary certificate for good and after sale service performance.
- 5. Eqpt. should be suitable to operate on BHEL power supply of 230V +-15%50Hz+-3% single phase AC supply and suitable voltage stablizer/CVT if recommended by the supplier should be supplied with the eqpt.
- 6. Supplier must quote all material /facility/service required for satisfactory commissioning /proving of eqpt will have to be arranged by the supplier.
- 7. Equipment should meet all general clauses enclosed and bidders must furnish point to point replies against all the points of our above specification and general clauses failing which their offer may not be considered.