

PORTABLE ANEMOMETER

Date-09/07/16

	NOTE:-	Vendors Remarks
1	The "VENDOR TO CONFIRM/SPECIFY" column of this format shall be filled in by the Vendor and submitted along with the offer. Inadequate / incomplete or ambiguous, information against any of the clauses of the specifications/requirements shall be treated as non-compliance.	
2	The offer and all documents enclosed with offer should be in English language only.	
3	Any deviation may please be clearly mentioned	
4	Vendor has to submit technical data sheets/documents of the model offered or a catalogue showing the product profile of the vendor.	
	SCOPE: SUPPLY & JOB PROVING OF PORTABLE ANEMOMETER COMPLYING WITH SPECIFICATIONS AS BELOW:	
Sl. NO.	CFFP/BHEL, HARDWAR REQUIREMENT	
1	Scope: The scope covers the supply & job proving of one no. portable Anemometer suitable for industrial application with necessary auxiliary accessories as per following specifications.	
2	Objective: For calculation of air intake (velocity and volume of flow) of different kind of combustion air blowers.	
3	Measurement Conditions	
3.1	Ambiant air temperature- (0-50° c)	
3.2	velocity range- 0.3-20 m/sec	
3.3	Accuracy- ±3%	
3.4	Resolution- 0.01m/sec	
4	Essential Accessories/ provisions :-	
4.1	Integral vane (100 mm)	
4.2	20 point averaging characteristics for min/max readings	
4.3	Flow rate	
4.4	Chargable battery(4-6 hrs of continuous operation) and a charger operating on 220volt (+10%) @ 50Hz	
	General Requirements :	
5	Documentation and manuals	
5.1	Two sets of documents (in soft and hard copy both) for detailed operation, maintenance and servicing manuals of equipment and other accessories to be provided in English language only.	

mmi
09/07/16
Prepared By:
Abhay Kumar
Engineer


Arjun Singh
09.07.16
Checked By:
Arjun Singh
Manager


Narendra Kumar
09/07/16
Approved By:
Narendra Kumar
DGM (M&S)


PORTABLE ANEMOMETER

Date-09/07/16

5.2	Calibration certificate must be provided along with instrument.	
6	Installation of instrument	
6.1	The supplier should take full responsibility for carry-out the job proving of instrument and all its accessories at BHEL free of cost. All capabilities of the instrument will have to be demonstrated to the satisfaction of the user. Training for handling and operation of the equipment should be provided to the user free of cost.	
7	Acceptance Norms	
7.1	10-15 sampling at CFFP BHEL haridwar.	
8	Guarantee :	
8.1	The supplier shall provide full warranty for at least two year from the date of job proving. If instrument remains out of order or starts malfunctioning the supplier shall be intimated of the same and the problem should be rectified within two weeks of getting intimated otherwise, the warranty period will be extended for the corresponding breakdown period .	


Prepared By:
Abhay Kumar
Engineer


Checked By:
Arjun Singh
Manager


Approved By:
Narendra Kumar
DGM (M&S)