

SPECIFICATION OF SETTLING TUBE WITH STAND, INDENT NO.20180796

ITEM NO.	MATERIAL CODE	MATERIAL DESCRIPTION	REQD. QTY.	BRIED SPECIFICATION
1.	QCXW06307957	Settling Tube with stand (Fluorescent)	05 NOS.	Centrifuge tubes should be pear-shaped, made from thoroughly annealed glass, and conform to the dimensions given in Figs. X5.1 of ASTM-E709-08. The graduations, numbered, should be clear and distinct.

Kamal Selgal
3/5/18
(E3, QC-T)


4/5/18
(E6A, QC-T)



