

BHARAT HEAVY ELECTRICALS LIMITED
Heavy Power Equipment Plant, Ramchandrapuram,
Hyderabad - 502 032. INDIA.

Enquiry No: _____
 Due Date: _____
 Supplier Qtn. No.: _____
 Date: _____

Ref No: CTP/TEST/SPECLAB/01
 Rev : 00
 Page: 01 of 02

TENDER DOCUMENTS TO BE FILLED BY THE SUPPLIER FOR SUPPLY OF
Instrument Storage and calibration workbench facility at Compressor test Plant instrument lab

ADDRESS OF THE SUPPLIER :

TELEPHONE NOS.:

e Mail Id:

FAX NOS.:

Sl. No.	Description for BHEL Requirement	Offered	Deviations	Remarks
1	Title			
	Instrument Storage and calibration workbench facility at Compressor test Plant instrument lab	Vendor to Conform		
2	Scope			
	The scope includes supply of lab furniture with storages, worktop, leg space/knee space, electrical switches and sockets, anti vibration tables.	Vendor to Conform		
3	Technical Specifications			
3.1	* Work Bench:			
	Table : 1500 width, 750 Height & 750 mm Depth 2 No's			
	Rear Side Panel 750 mm Depth & 750 mm Height			
	450mm Width, 450 mm Depth, 506 mm Height Metallic Drawers (Storage Racks) 1S, 1D : 4 No's			
	Leg Space 600 mm Width and 750 mm Height			
	Filler Panel 750mm Height 2 No's			
	Table : 600 width, 750 Height & 750 mm Depth 4 No's			
	600mm Width, 450 mm Depth, 506 mm Height Storage Racks 2S, 1D : 4 No's			
	Pipe Frame of 750 Height & 750 mm Depth			
	CRCA Metal or Equivalent Sheet Thickness Minimum of 1.5 mm for Table			
	* Refer Attached Drawing for more information.			

B. Sivabala

BK
 108/RS/114 RTG
 108/RS/114 RTG
 DGM / Testing (Compressor & DT)
 BHEL, HYD-32

3.2	Accessories:			
	a) Table Mounted Electrical Trunking: 06 No's with the following			
	Electrical Sockets: 15/ 5A Electrical Plug Sockets with switch control for each Work bench			
	b) Anti Vibration Pads of 15 mm thickness			
3.3	Work Top:			
	Material:			
	1) Granite or			
	2) ASS TOP - Staron Aspen Snow AS610, Make - Samsung			
	(18mm thk. MDF + 12mm thk. ASS Top)			
3.4	Paint: Powder Coating shall be of minimum 50 microns thickness.			
4	Documentation			
	Dimensional drawings/details & Catalogues shall be furnished for approval before manufacturing and 2 sets of O&M Manuals along with the final drawings shall be furnished along with equipment.	Vendor To Conform		
5	Packing and forwarding			
	Dispatch shall be done in closed packing boxes consistent requirements of safe transit. It shall be the supplier responsibility to determine that the package done is adequate to preserve integration of the unit after being met the satisfactory visual check and such that the complete unit shall arrive at its destination in an undamaged condition (No Mechanical damages, distortions dust and rain water entry etc.) and the unit is ready for its intended use.	Vendor To Conform		
6	Pre Qualifying Criteria			
6.1	Vendor Shall be OEM or otherised agent of similar equipment	Vendor To Conform		
6.2	Vendor Should have supplied at least two similar equipments in the last three years	Vendor To Conform		
6.3	Vendor shall submit along with the offer the list of customers where the similar type of equipment have been supplied	Vendor to Provide details		
7	Delivery Period (Vendor to mention)	Vendor To Conform		

B. Komala Rao

B. Komala Rao

B. Komala Rao

उप महा प्रबंधक / परीक्षण (कंप्रेसर डी टी)

DGM / Testing (Compressor & DT)

बी.एस.ई.एल. हैदराबाद, BHEL, HYD-32