



BHARAT HEAVY ELECTRICALS LIMITED,
RAMACHANDRAPURAM, HYDERABAD :: 32


**SPECIFICATION
FOR**

API REFERENCE MASTER ROTARY SHOULDERED CONNECTION GAUGE 7 5/8" LEFT HAND SET (PLUG & RING)

Ref. no. HY / QC / CGC / API 7-2

DATE 21/10/2019

Annexure 'A'

SL	DESCRIPTION OF BHEL REQUIREMENT	SPECIFIED / TO BE CONFIRMED BY	OFFERED	DEVIATIONS	REMARKS
1.0	PURPOSE OF THE GAUGE SET For Calibration of API Rotary Shouldered Connection Working Gauges size 7 5/8" Left Hand set(Plug & Ring) used in BHEL, Ramachandrapuram, Hyderabad, India.	vendor to confirm			
2.0	SCOPE Supply of API reference master rotary shouldered connection gauges 7 5/8" Left Hand set (Plug & Ring) with Calibration certificate, Calibrated at API approved metrology Institutes like NIST-USA, NIM-CHINA etc. (API APPROVED CERTIFICATE TO BE ENCLOSED)	vendor to confirm			
3.0	SPECIFICATION : API SPEC 7-2 (Latest Edition)	vendor to confirm			
4.0	CARRYING AND TRANSPORTATION Water, Dust, Shock proof carrying case.	vendor to confirm			
5.0	COMMISSIONING Supplier shall commission the above Gauge set at Bharat Heavy Electricals Limited (BHEL), Ramachandrapuram, Hyderabad-502 032, India.	Vendor to confirm			
6.0	DOCUMENTATION: Following documents shall be supplied by the vendor				
6.1	Operating instruction manual in English, 2 copies	Vendor to confirm			
6.2	Calibration Certificate of API reference master 7 5/8" Left Hand gauge set (Plug & Ring) Calibrated at API approved metrology Institutes like NIST-USA, NIM-CHINA etc.	Vendor to confirm			
7.0	QUALIFYING CONDITIONS / REFERENCE LIST				
7.1	Only those vendors who have supplied and commissioned at least one such gauge set and it has been working satisfactorily.	Vendor to confirm			
7.2	Name of the customer where referred gauge is supplied with purchase order no. and date.	Vendor to Furnish.			
 KOTHA MAHESH Sr. Engineer/Quality Control-SMC.CCG BHEL - HPEP, HYD-32  M. Jyoti Basu DGM / Quality Control BHEL, HYD-32					