

Pre-Qualifying Conditions for Tender No.- C/7029/11/4327T

1. The Manufacturer should have standing of manufacturing the equipment for minimum 5 years.
2. The Manufacturer should furnish the list of customers where same model of equipment has been commissioned in last 5 years from the date of opening of the Tender.
3. The vendor should furnish satisfactory performance certificate from one customer where same model has been installed and has been working satisfactorily for last at least 1 year from the date of opening of the Tender. Also furnish the details of customer/company where the equipment has been installed with contact address and telephone no. of the customer.
4. The vendor should ensure regular after sales service including spares for at least five years after expiring of Warranty period / *Guarantee period*

*S.K.G.*

*S.K.G.*

*27/3/12*

Dr. S.K. Gupta  
AGM (Surgery)

9

0

**AUTOREFRACTOMETER**

**Technical Specifications**

**Requirements:**

1. Autorefractometer with standard accessories and Motorised-table

**Specifications:**

**Measurement Range:** Sphere -30.00 to +25.00 D

Cylinder 0 to +/- 12.00 D

Axis 0 to 180

**Measurable min Pupil diameter:** 2 mm

**Auto shooting**

**Chart:** Scenery chart

**PD measurement range:** 30 to 85 mm


**Display:** Tilttable 5.7 inch color LCD

**Printer:** Built in thermal type line printer

**Accessories:** Interface Cable, Barcode Scanners, Model eye

**Table:** Motorised

**Warranty/ Guarranty** 1 year may be asked from the supplier from the date of commissioning.

  
(Dr. S. Das)  
19-01-2012  
Sr. Consultant (Ophth.)