

TECHNICAL SPECIFICATIONS- DIGITAL CENTRIFUGE

- 1- Digital centrifuge with digital speed meter and brushless induction motor.
- 2- Speed regulator facility with Centrifuge speed range 1000 rpm to 5000 rpm or beyond.
- 3- Timer facility with range 1 minute to 15 minutes or beyond.
- 4- Brake facility.
- 5- imbalance detector facility.
- 6- lid lock facility.
- 7- rotor head capacity 16x15 ml.
- 8- Service center in Hyderabad. Vendor has to provide service center address along with technical bid.
- 9- Vendor should be able to demonstrate the equipment at BHEL General Hospital, Hyderabad, if asked for.
- 10- Pre-Qualification Criteria: Vendor has to submit customer reference list (minimum two customers) using the same equipment, with their contact details (phone number of customers must be provided).